



Nederlands Meetinstituut

Member State

The Netherlands

OIML Certificate N°R76/1992-NL-95.30

Revision 2

Project number 10067126

Page 1 of 2

OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name: NMI Certin B.V.
Address: Hugo de Grootplein 1, Dordrecht
Person responsible: A.J. Nederlof

Applicant

Name: A&D Instruments Ltd.
Address: Abingdon science Park
Abingdon, Oxford OX14 3YS
United Kingdom

Manufacturer of the certified pattern

Name: A&D Instruments Ltd.
Address: Abingdon science Park
Abingdon, Oxford OX14 3YS
United Kingdom

Identification of the certified pattern

Type: HR-EC

Max \leq 210 g or \leq 1050 ct

e \geq 1 mg or \geq 0.005 ct

n \leq 210000 divisions

e = d or e = 10 d

Temperature limit 15 °C / 25 °C

This certificate attests the conformity of the above-mentioned pattern (represented by the samples identified in the associated test report, the EC type-approval certificate and the description with number T2733 and the appertaining documentation folder) with the requirements of the following Recommendation(s) of the International Organization of Legal Metrology (OIML):

R76
edition 1992
for accuracy class **I**

Nederlands Meetinstituut
Hugo de Grootplein 1
3314 EG Dordrecht (NL)
Telephone +31 78 6332332
Telefax +31 78 6332309

NMI B.V. (Registered at the Chamber of Commerce
Delft number 26028701)

Subsidiary companies:
NMI Certin B.V. (26033418)
NMI Van Swinden Laboratorium B.V. (26028703)
NMI Inspecties en Kansspeltechniek B.V. (26028700)
NMI International B.V. (26039176)

This certificate is issued under the provision that
NMI B.V. nor its subsidiary companies accept any
liability.

Reproduction of the complete certificate is allowed.
Parts of the certificate may only be reproduced
after written permission.



**QUALIFIED
BY STERLAB**
Reg. nr. L 029



Nederlands Meetinstituut

Member State

The Netherlands

OIML Certificate N°R76/1992-NL-95.30

Revision 2

Project number 10067126

Page 2 of 2

This certificate relates only to the metrological and technical characteristics of the pattern of the instrument concerned, as covered by the relevant OIML International Recommendation(s).

This certificate does not bestow any form of legal international approval.

The conformity was established by tests described in the associated test report N° R76/1992-NL-95.30, that includes 38 pages.

This revision replaces the earlier version(s) of the OIML certificate of conformity with number R76/1992-NL-95.30.

The issuing authority
A.J. Nederlof

9 December 1996

The OIML member
G.J. Faber

9 December 1996

*
* *

Important note: Apart from the mention of the certificate's reference number and the name of the OIML Member State in which the certificate was issued, partial quotation of the certificate or of the associated test report is not permitted, though they may be reproduced in full.