

OIML Member State
The Netherlands

Number R76/2006-A-NL1-20.01
Project number 2379186
Page 1 of 2

Issuing authority NMI Certin B.V.
Person responsible: F. van Booma – de Smit

Applicant and Manufacturer A&D Instruments Ltd.
Unit 24/26 Blacklands Way, Abingdon Business Park
Abingdon, Oxfordshire, OX14 1DY
United Kingdom

Identification of the certified type **A Terminal**
Type : AD-4329A-DLC

Characteristics See next page

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76 - Edition 2006 for accuracy class $\text{\textcircled{III}}$ or $\text{\textcircled{IIII}}$

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMI Certin B.V., OIML Issuing Authority NL1**
7 January 2020

Certification Board

NMI Certin B.V.
Thijsseweg 11
2629 JA Delft
The Netherlands
T +31 88 6362332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMI Certin B.V. as Issuing Authority can be verified at www.oiml.org

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon at the top of the electronic version of this certificate.



The conformity was established by the results of tests and examinations provided in the associated OIML Reports:

- No. NMI-2379186-01 dated 7 January 2020 that includes 17 pages;
- No. NMI-2379186-02 dated 7 January 2020 that includes 18 pages.

Characteristics of the Terminal:

Configuration	Digital load cells
Accuracy class	OIML R 76
Weighing range(s)	III or IIII
Maximum number of scale intervals	Single interval Multi-interval
Maximum number of weighing ranges	-
Load cell power supply voltage	1
Fraction of the maximum permissible error	8 V DC
Temperature range	0
Power supply voltage	-10 °C / +40 °C
Software identification	220 – 240 V AC 50/60 Hz
	Version number: r 1xx (xx is a number between 00 and 99 and represents the non-legally relevant software)