

Issued by NMI Certin B.V.  
Hugo de Grootplein 1  
3314 EG Dordrecht  
The Netherlands

Notified Body Number 122

In accordance with The Council Directive 90/384/EEC on Non-Automatic Weighing Instruments.

Applicant A&D Instruments Ltd.  
Abingdon Science Park  
Abingdon, Oxford OX14 3YS  
United Kingdom

In respect of A Class **II**, graduated, self-indicating electronic single-interval **non-automatic weighing instrument**.  
Manufacturer : A&D  
Type : HF and HF-G

Characteristics  $n \leq 41000$  divisions  
 $\text{Max} \leq 8100 \text{ g}$  or  $\leq 40500 \text{ ct}$   
 $e \geq 10 \text{ mg}$  or  $\geq 0.05 \text{ ct}$   
 $e = 10d$   
Temperature limit:  $10 \text{ }^\circ\text{C} / 30 \text{ }^\circ\text{C}$


In the description T2699 Revision 4 further essential characteristics are described.

Valid until 23 October 2005

Description and documentation The weighing instrument is described in the description number T2699 Revision 4 and documented in the documentation folder(s) number T2699-1, appertaining to this EC type-approval certificate.

Remarks This revision EC type-approval certificate replaces the earlier versions, except for its documentation folder.

Dordrecht, 9 December 1996  
NMI Certin B.V.



A.J. Nederlof  
Director

**Nederlands Meetinstituut**  
**Hugo de Grootplein 1**  
**3314 EG Dordrecht (NL)**  
Telephone +31786332332  
Telefax +31 78 6332309

Nederlands Meetinstituut NY (Registered at the Chamber of Commerce Delft number 28701)

Subsidiary companies:  
NMI Cernn BV (33418)  
NMI Van Swinden Laboratorium B.V. (28703)  
NMI Inspecties & Kansspeltechniek B.V.  
(28700) NMI International B.V. (39176)

This certificate is issued under the provision that Nederlands Meetinstituut N.V nor its subsidiary companies accept any liability

Reproduction of the complete certificate is allowed Parts of the certificate may only be reproduced after written permission.



QUAUFIED  
BY STERLAB  
Reg. nr. L 029