

Number **T2884** Revision 2  
Project number 10079265 Page  
1 of 4

Issued by NMI Certin B.V.  
Hugo de Grootplein 1  
3314 EG Dordrecht  
The Netherlands

Notified Body Number 122

In accordance with The Council Directive 90/384/EEC on Non-Automatic Weighing Instruments.

Applicant A&D Instruments Ltd.  
Abingdon Science Park  
Abingdon Oxford  
OX143YS  
United Kingdom

In respect of A Class **I**, graduated, self-indicating, electronic, single-interval, **non-automatic weighing instrument**.  
Manufacturer : A&D  
Type : HM

Characteristics  $n \leq 310000$  divisions  
 $\text{Max} \leq 310 \text{ g}$  or  $\text{Max} \leq 1550 \text{ ct}$   
 $e \geq 1 \text{ mg}$  or  $e \geq 0.005 \text{ ct}$   
 $e = d$  or  $e = 10 d$  or  $e = 100 d$   
Temperature limit:  $+10 \text{ }^\circ\text{C}$   $\cdot/\cdot$   $+30 \text{ }^\circ\text{C}$

In the description T2884 Revision 2 further characteristics are described.

Valid until 17 October 2006

Description and documentation The weighing instrument is described in the description number T2884 Revision 2 and documented in the documentation folder number T2884-1, appertaining to this EC type-approval certificate.

Remarks This revision EC type-approval certificate replaces the earlier version(s), except for its documentation folder.

Dordrecht, 10 November 1997

NMI Certin B.V.



A.J. Nederlof  
Director

Nederlands Meetinstituut  
Hugo de Grootplein 1  
3314 EG Dordrecht (NL)  
Telephone +31 786332332  
Telefax +31 786332309

NMI BV (Chamber of Commerce  
Haaglanden No 27228701)

Subsidiary companies:  
NMI Certin B.V. (27233418)  
NMI Van Swinden laboratorium B.V. (27228703) NMI  
Inspecties en Kansspeltechniek B.V. (27228700)  
NMI International BV. (27239176)

This certificate is issued under the provision that NMI B V  
nor its subsidiary companies accept any liability

Reproduction of the complete certificate is allowed Parts of  
the certificate may only be reproduced after written  
permission



QUALIFIED  
BY STERIAB  
Reg. hr. L 029