

Issued by NMI Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands

Notified body number 122

In accordance with Paragraph 8.1 of the European Standard on Metrological aspects of non-automatic weighing instruments EN 45501:1992/AC:1993. The applied error fraction P_i , meant in paragraph 3.5.4. of this standard is 0.5.

Applicant A&D Instruments Ltd.
Abingdon Science Park
Abingdon, Oxford, OX14 3YS
United Kingdom

In respect of The model of an **indicator**, tested as part of a weighing instrument (for non- automatic weighing instrument class **III** and **III**).
Manufacturer : A&D
Type : HV-D

Characteristics Electronic, self-indicating device, with single-interval indication.
The maximum number of verification scale intervals will be:
 $n \leq 3000$ for class **III** instruments or
 $n \leq 1000$ for class **III** instruments.
In the description TC2880 Revision 1 further characteristics are described.

Description and documentation The indicator is described in the description number TC2880 Revision 2 and documented in the documentation folder number TC2880-1, appertaining to this test certificate.

Remarks Summary of the tests involved: see Appendix number TC2880 Revision 2. This revision test certificate replaces the earlier versions except for its documentation folder.

Dordrecht, 16 December 1996
NMI Certin B.V.



A.J. Nederlof
Director